

Asselum SL

Headquarters
 Polig.Ind. Can Roqueta c/ de Ca n'Alsina, 76
 08202 Sabadell (Barcelona) – España
 Tel +34-93-7259810 -
 e-mail info@asselum.com
 Web: www.asselum.com

Palazzoli SpA Laboratory Surveillance Certificate

| | | | |
|-------------------|--------------------------|-------------|--|
| Report n. | 1008g-02/CL1738/001/2025 | Date | July 15th 2025 |
| Customer | Palazzoli SpA | | |
| Address | Via Palazzoli, 31 | Postal code | 25128 |
| Place | Brescia | Province | Brescia |
| State | Italy | | |
| Contact | Ing. Massimo Legrenzi | e-mail | massimo.legrenzi@palazzoli.it |
| Phone | +39-030-20151 | | |
| Alternate contact | Sig. Ivo Meroni | e-mail | ivo.meroni@palazzoli.it |
| Phone | +39-030-2015146 | | |

Summary

| | | |
|---|--|---|
| 1 | Audit..... | 2 |
| 2 | Scope of the Laboratory Surveillance Certificate | 2 |
| 3 | Period of Validity of the Surveillance..... | 2 |
| 4 | Contractual Rules | 2 |
| 5 | Approved by..... | 2 |

1 Audit

The structures and the test laboratory personnel of the aforesaid manufacturer have been evaluated in accordance with the following standards:

- ▶ UNI EN 13032-1:2012 – Measurement and presentation of photometric data of lamps and luminaires, Part: 1 Measurement and file format
- ▶ UNI EN 13032-4:2015 – Measurement and presentation of photometric data of lamps and luminaires, Part: 4 LED lamps, modules and luminaires

2 Scope of the Laboratory Surveillance Certificate

Keep the photometric measurements of LED lamps and luminaires made according to the following international standards or recommendations under surveillance:

- ▶ UNI EN 13032-1: 2012
- ▶ UNI EN 13032-4: 2015
- ▶ CIE S-025/E:2015
- ▶ IES LM-79-19

that means, test methods for the measurement of the total luminous flux and the photometric characterization of LED modules.

3 Period of Validity of the Surveillance

| | |
|-----------------|--------------------|
| Date of issue | July 1st 2025 |
| Expiration date | December 31th 2025 |

4 Contractual Rules

- ▶ The validity of this authorization for surveillance is maintained with subsequent audit at the laboratory, to be agreed with the company subject of the authorization
- ▶ The Surveillance Certificate can be reproduced in full in technical documentation, marketing documentation or on the company website with the written authorization of the company that manages the surveillance
- ▶ The Surveillance Certificate can be copied and transferred to third parties for scopes related to public tenders where it is specifically requested
- ▶ The Surveillance Certificate is not transferable to third parties for the same scope
- ▶ It is not allowed to provide test reports derived from measurements performed at the laboratory in which the trademark or the reference to the accreditation of the company managing the surveillance is explicitly indicated

5 Approved by

| | |
|---|---|
| Asselum SL ENAC 1321/LE2474 |  |
| Borja LLuch Mencucci (Laboratory technician) | Marc Ballbè (CEO) |